ΑD)

Award Number: DAMD17-99-1-9239

TITLE: Prediction of Pathologic Fracture Risk in Activities of Daily Living and Rehabilitation of Patients with Metastatic Breast Carcinoma of the Pelvis and Femur

PRINCIPAL INVESTIGATOR: Nozomu Inoue, M.D., Ph.D.

CONTRACTING ORGANIZATION: The Johns Hopkins University

Baltimore, Maryland 21205-2196

REPORT DATE: August 2001

TYPE OF REPORT: Annual

PREPARED FOR: U.S. Army Medical Research and Materiel Command

Fort Detrick, Maryland 21702-5012

DISTRIBUTION STATEMENT: Approved for Public Release;

Distribution Unlimited

The views, opinions and/or findings contained in this report are those of the author(s) and should not be construed as an official Department of the Army position, policy or decision unless so designated by other documentation.

REPORT DOCUMENTATION PAGE

Form Approved OMB No. 074-0188

Public reporting burden for this collection of information is estimated to average 1 hour per response, including the time for reviewing instructions, searching existing data sources, gathering and maintaining the data needed, and completing and reviewing this collection of information. Send comments regarding this burden estimate or any other aspect of this collection of information, including suggestions for reducing this burden to Washington Headquarters Services, Directorate for Information Operations and Reports, 1215 Jefferson Davis Highway, Suite 1204, Arlington, VA 22202-4302, and to the Office of

Management and Budget, Paperwork Reduction Project (0704-0188), Washington, DC 20503							
1. AGENCY USE ONLY (Leave blank) 2. REPORT DATE 3. REPORT TYPE AND			DATES COVERED				
	August 2001	Annual (15 Jul					
4. TITLE AND SUBTITLE			5. FUNDING NUN				
Prediction of Pathologic	DAMD17-99-1	-9239					
Daily Living and Rehabil							
Metastatic Breast Carcin							
6. AUTHOR(S)							
Nozomu Inoue, M.D., Ph.D.							
7. PERFORMING ORGANIZATION NAM	ME(S) AND ADDRESS(ES)		8. PERFORMING	ORGANIZATION	_		
7. TEM ORIGINA ORGANIZATION NAME	ME(O) AND ADDITION(LO)		REPORT NUMB				
The Johns Hopkins Univer	sity						
Baltimore, Maryland 212							
-							
E-Mail: ninoue@jhmi.edu							
9. SPONSORING / MONITORING AGE	NCY NAME(S) AND ADDRESS(ES	;)	10. SPONSORING	/ MONITORING			
			AGENCY REP	ORT NUMBER			
U.S. Army Medical Research and M	Nateriel Command						
Fort Detrick, Maryland 21702-501	2		ļ				
			20101	200			
11. SUPPLEMENTARY NOTES		/IIII	20124	/IIV			
Report contains color		LUUI					

12a. DISTRIBUTION / AVAILABILITY STATEMENT

Approved for Public Release; Distribution Unlimited

12b. DISTRIBUTION CODE

13. ABSTRACT (Maximum 200 Words)

The purpose of the project is to develop a computer model of the pelvis and proximal femur which can be used to predict pathologic fracture risk and study the effects of pelvic and proximal femoral metastatic bone lesions on the care and management of breast cancer patients. The scope of the research is to construct graphical and quantitative models of the pelvis and proximal femur on a computer workstations including Finite Element Method (FEM) and Discrete Element Method (DEM) to study the stress and strain in the pelvis and proximal femur and pressure distribution of the hip joint in the patient with metastatic bone lesions of the breast cancer in the pelvis and proximal femur with interactive capability. A simple FEM model with a bone defect was analyzed and the results were validated by mechanical testing. The DEM model was used to estimate the pressure distribution of the hip joint during Activity in Daily Living (ADL), including walking, climbing stairs, and getting up from a chair.

14. SUBJECT TERMS Breast Cancer, Bone Met Reality Computer Model,	15. NUMBER OF PAGES 14 16. PRICE CODE		
17. SECURITY CLASSIFICATION OF REPORT Unclassified	18. SECURITY CLASSIFICATION OF THIS PAGE Unclassified	19. SECURITY CLASSIFICATION OF ABSTRACT Unclassified	20. LIMITATION OF ABSTRACT Unlimited

Table of Contents

Cover	1
SF 298	2
Table of Contents	3
Introduction	4
Body	4
Key Research Accomplishments	6
Reportable Outcomes	6
Conclusions	6
References	7
Figures	9
Annendix 1	13

INTRODUCTION

The subject of the project is to develop models to predict the pathological fracture risk in activities in daily living life (ADL), nursing care, and rehabilitation in breast cancer patients with metastatic lesion in the pelvis and proximal femur. The purpose of the project is to develop a computer model of the pelvis and proximal femur which can be used to predict the pathologic fracture risk and study the effects of pelvic and proximal femoral metastatic bone lesions on the care and management of breast cancer patients. The scope of the research is to include the construction of graphical and quantitative models of the pelvis and proximal femur on computer workstations including Finite Element Method (FEM) and Discrete Element Method (DEM) to study the stress/strain in the pelvis and proximal femur and pressure distribution of the hip joint in the patient with metastatic bone lesions of the breast cancer in the pelvis and proximal femur with interactive capability.

BODY

This section describes the research accomplishment associated with each Task outlined in the approved Statement of Work (written in italic letters).

Technical Objectives 1: Computer model construction

Task 1: Months 1-6: Establishment of database for location, size, and distribution of metastatic breast cancer to pelvic and femoral regions.

Completed as described in the first annual report.

Task 2: Months 6-18: Development of Multi-Discrete Element Model (DEM) and Finite Element Model (FEM) of pelvic and femoral regions based on the database established in Task 1.

An interactive DEM model of the hip joint has been developed as described in the first annual report. As described in Task 6, the loading conditions during the activities daily living (ADL) are now incorporated in the model. A simplified FEM model of the long bone with a bone defect was created (Fig. 1). The results of the FEM was validated by actual mechanical testing (see Task 4). The pelvic FEM model is being developed.

Task 3: Months 6-12: Mechanical testing using cadaveric specimens of pelvic region with and without bone defects in the pelvis.

See Task 4.

Task 4: Months 13-18: Mechanical testing using cadaveric specimens of pelvic region with and without bone defects in the proximal femur.

Both the Dynamic Joint Simulator and MTS BIONIX 858 testing machine are now functioning well at Ross facility. An Acoustic Emission (AE) technique was applied to monitor the microfracture of the cadaveric bone specimen with bone defect during mechanical testing

(Figs. 2 and 3). Mechanical testing was performed on the long bone shaft to validate the results of the FEM model (see Task 2). Fractures were initiated from the corners of the defect under torsional testing where the highest stress/strain was calculated in the FEM model (Fig. 1).

In the original protocol of mechanical testing, we planned to apply loading to the central part of the hip joint. However, the highest pressure in the hip joint was estimated at the posterior part of the hip joint during some of activities in daily living as described in Task 6, and this area appears to be critical for pathologic fractures. Therefore, it is important to identify to the most critical loading condition and apply this loading condition for mechanical testing of the pelvis and proximal femur with bone defect. Detailed loading conditions will be estimated by Tasks 6 and 7; for this reason, Tasks 3 and 4 will be continued in year 3 parallel to Tasks 6 and 7.

Technical Objectives 2: Establishment of the model to predict fracture risk in activities in daily living life, nursing care, and rehabilitation

Task 5: Months 19-22 Acquisition of kinematic and force data in Activity in Daily Living (ADL), rehabilitation program, and nursing care.

See Task 6.

Task 6: Months 23-28: Analysis of the loading conditions during the activities studied in Task 5.

The loading conditions during activities in daily living, including walking, climbing stairs, and chair-up, were obtained and incorporated in the DEM analysis. The contact pressure distribution during the activities has been analyzed on the hip joint without the defect. The preliminary results of the analysis is presented in Fig. 4 (walking), Fig. 5 (climbing stairs), and Fig. 6 (chair-up).

The following tasks will be performed after Month 23 as planed.

Task 7: Months 29-33: Analysis of the stability and stress/strain distribution in the metastatic pelvis and proximal femur under the loading conditions predicted in Task 6.

Task 8: Months 34-36: Preparation of publications.

In addition to the approved Tasks, we have established a virtual interactive model of external fixation of a long bone. This simulation and visualization model could be expanded to simulate fixation of the pelvic fracture. This study has been presented at ASME-BED (American Society of Mechanical Engineering, Bioengineering Division) 2001 Bioengineering Conference, Snowbird, Utah, June 27-July 1, 2001 (Appendix 1).

KEY RESEARCH ACCOMPLISHMENT

- The basis of the biomechanical model using FEM to estimate the stress/strain distribution around the bone defect has been developed and validated by the mechanical testing.
- Pressure distribution of the hip joint during walking, climbing stairs, and chair-up was estimated by DEM analysis.

REPORTABLE OUTCOMES

• Manuscripts, abstracts, presentations;

Abstract:

Kim YH, Elias JJ, Inoue N, Chao EYS: Study on external fixator adjustability using simulation and visualization model. Proceeding of 2001 Bioengineering Conference, American Society of Mechanical Engineering (ASME), BED-Vol. 50: 237-238, 2001.

Presentation

Inoue N, Chao EYS. Computer-aided preoperative planning for acetabular osteotomy and total hip arthroplasty. 1st International Symposium and Workshop on Reconstructive Surgery of the Pelvis. Graz, Austria, April 18-20, 2001.

Patents and licenses applied for and/or issued;

None

Degrees obtained that are supported by this award;

None

Development of cell lines, tissue or serum repositories;

None

• Informatics such as databases and animal models, etc;

None

Funding applied for based on work supported by this award;

None

• Employment or research opportunities applied for and/or received on experiences/training supported by this award.

None

CONCLUSIONS

A simple finite element model with a bone defect was analyzed and the results were validated by mechanical testing. The discrete element model was used to estimate the pressure distribution of the hip joint during Activity in Daily Living (ADL), including walking, climbing stairs, and getting up from a chair.

REFERENCES

(1) Refereed Journal Articles

Elias JJ, Nagao M, Chu YH, Carbone JJ, Lennox DW, Chao EY: Medial cortex strain distribution during noncemented total hip arthroplasty. Clin Orthop 2000 Jan; (370):250-8.

Shermak MA, Wong L, Inoue N, Nicol T: Reconstruction of complex cranial wounds with demineralized bone matrix and bilayer artificial skin. J Craniofac Surg 2000 May;11(3):224-31.

Jani MM, Sponseller PD, Gearhart JP, Barrance PJ, Genda E, Chao EY: The hip in adults with classic bladder exstrophy: a biomechanical analysis. J Pediatr Orthop 2000 May-Jun;20(3):296-301.

Shin DS, Choong PF, Chao EY, Sim FH: Large tumor endoprostheses and extracortical bone-bridging: 28 patients followed 10-20 years. Acta Orthop Scand 2000 Jun;71(3):305-11.

Meffert RH, Tis JE, Inoue N, McCarthy EF, Brug E, Chao EY: Primary resective shortening followed by distraction osteogenesis for limb reconstruction: a comparison with simple lengthening. J Orthop Res 2000 Jul;18(4):629-36.

Inoue YZ, Frassica FJ, Sim FH, Unni KK, Petersen IA, McLeod RA: Clinicopathologic features and treatment of postirradiation sarcoma of bone and soft tissue. J Surg Oncol 2000 Sep;75(1):42-50.

Chu Y, Elias JJ, Duda GN, Frassica FJ, Chao EY: Stress and micromotion in the taper lock joint of a modular segmental bone replacement prosthesis. J Biomech 2000 Sep;33(9):1175-9.

Frassica DA, Thurman S, Welsh J: Radiation therapy. Orthop Clin North Am 2000 Oct;31(4):557-66.

Pierce MC, Valdevit A, Anderson L, Inoue N, Hauser DL: Biomechanical evaluation of dual-energy X-ray absorptiometry for predicting fracture loads of the infant femur for injury investigation: an in vitro porcine model. J Orthop Trauma 2000 Nov;14(8):571-6.

Elias JJ, Frassica FJ, Chao EY: The open section effect in a long bone with a longitudinal defect - a theoretical modeling study. J Biomech 2000 Nov;33(11):1517-22.

Frassica FJ, Khanna JA, McCarthy EF: The role of MR imaging in soft tissue tumor evaluation: perspective of the orthopedic oncologist and musculoskeletal pathologist. Magn Reson Imaging Clin N Am 2000 Nov;8(4):915-27.

Virolainen P, Inoue N, Young DR, Frassica FJ, Chao EY: Method to remove intramedullary cement prior to mechanical testing in canine segmental replacement model. Biomed Mater Eng 2000;10(2):51-6.

Sparks AB, Peterson SN, Bell C, Loftus BJ, Hocking L, Cahill DP, Frassica FJ, Streeten EA, Levine MA, Fraser CM, Adams MD, Broder S, Venter JC, Kinzler KW, Vogelstein B, Ralston SH: Mutation screening of the TNFRSF11A gene encoding receptor activator of NF kappa B

(RANK) in familial and sporadic Paget's disease of bone and osteosarcoma. Calcif Tissue Int 2001 Mar;68(3):151-5.

Bhimani MA, Wenz JF, Frassica FJ: Pigmented villonodular synovitis: keys to early diagnosis. Clin Orthop 2001 May;(386):197-202.

Rose PS, Lietman SA, McCarthy EF, Frassica FJ: Destructive scapular lesion in an infant. Clin Orthop 2001 May; (386):260-2, 264-7.

Fung M, Kato S, Barrance PJ, Elias JJ, McFarland EG, Nobuhara K, Chao EY: Scapular and clavicular kinematics during humeral elevation: A study with cadavers. J Shoulder Elbow Surg 2001 May-Jun;10(3):278-85.

Genda E, Iwasaki N, Li G, MacWilliams BA, Barrance PJ, Chao EY: Normal hip joint contact pressure distribution in single-leg standing-effect of gender and anatomic parameters. J Biomech 2001 Jul;34(7):895-905.

Larsson S, Kim W, Caja VL, Egger EL, Inoue N, Chao EY: Effect of early axial dynamization on tibial bone healing: a study in dogs. Clin Orthop 2001 Jul;388:240-51

Mont MA, Jones LC, Elias JJ, Inoue N, Yoon TR, Chao EY, Hungerford DS: Strut-autografting with and without osteogenic protein-1: a preliminary study of a canine femoral head defect model. J Bone Joint Surg Am 2001 Jul;83-A(7):1013-22

(2) Abstracts

Kim YH, Elias JJ, Inoue N, Chao EYS: Study on external fixator adjustability using simulation and visualization model. Proceeding of 2001 Bioengineering Conference, American Society of Mechanical Engineering (ASME), BED-Vol. 50: 237-238, 2001.

(3) Book Chapter

Chao EY, Inoue N, Elias JJ, Frassica FJ: Image-based computational biomechanics of the musculoskeletal system. In: Bankman, ed. *Handbook of Medical Imaging*. Chap. 18, PP 285-298. The Netherlands: Academic Press.



Fig. 1 Strain energy density (SED) distribution superimposed over a finite element (FE) model with a defect for a torsional load.




Fig. 2 Torsional failure test of the diaphysis including the defect. An acoustic emission (AE) sensor was attached to the test specimen to detect microfractures during testing.




Fig. 3 Torque-torsional angle and AE accumulation-torsional angle curves.




Fig. 4 Showing the contact distribution of the acetabulum during walking of four phases: Heel-Contact, Midstance, Toe-Off, and Swing Phase (% of cycle).




Fig. 5 Showing the contact distribution of the acetabulum during climbing stairs of four phases (% of cycle).



Fig. 6 Showing the contact distribution of the acetabulum during chair up of four phases: Sitting, Hip off, Inclining, Standing (% of cycle).

STUDY ON EXTERNAL FIXATOR ADJUSTABILITY USING SIMULATION AND VISUALIZATION MODEL

Yoon Hyuk Kim, John J. Elias, Nozomu Inoue and Edmund Y.S. Chao

Orthopaedic Biomechanics Laboratory
Johns Hopkins University
Baltimore, Maryland

INTRODUCTION

An external fixator is commonly used to stabilize fractured long bone segments. After inserting the pins and applying the fixator, adjustment of the position of each fragment is often necessary to reduce the fracture site and correct any residual deformities. A primary goal of external fixator design is allowing adequate adjustability for fracture reduction. The combination of an external fixator and a fractured bone can be modeled as a multi-link kinematic mechanism. Using computational techniques, the translations and rotations in the connecting joints of the mechanism necessary to reduce a given fracture can be rapidly quantified. Applying the calculated translations and rotations to a virtual biomechanical model of the fixator and bone system allows 3-D visualization of the adjustments required to reduce a fracture. The objective of this paper is to present a computational and graphical model of an external fixator that can be used to evaluate the fixator design, optimize the design, and guide clinicians using the device.

METHODS

The simulation model was based on an Orthofix, D.A.F external fixator. The model includes the mathematical representation of the motion of each link of the fixator and the bone segments as a kinematic linkage connected by various joints (Fig. 1). The distal bone fragment (D) was fixed to the origin of the global coordinate system (G), with the primary axes of the bone aligned with the global axes. The position of the proximal bone (P) within the global coordinate system was described using a 4×4 homogeneous transformation matrix, ${}^{G}\mathbf{T}_{P}$, which can be determined radiographically using established bone landmarks [1, 2]. This matrix is equivalent to the transformation of each link of the fixator from the proximal bone segment to the distal bone within the global coordinate system, as:

$${}^{G}\mathbf{T}_{P} = {}^{G}\mathbf{T}_{D} \cdot {}^{D}\mathbf{T}_{1} \cdot {}^{1}\mathbf{T}_{2} \cdot {}^{2}\mathbf{T}_{3} \cdot {}^{3}\mathbf{T}_{4} \cdot {}^{4}\mathbf{T}_{P},$$
 (1),

where ${}^{G}T_{D}$ is a 4 × 4 unit matrix representing the position of the distal bone in the global coordinate system, and the other transformations represent transformation of each segment of the linkage (Fig. 1). ${}^{D}T_{1}$

and 4T_P represent rigid body translations in the directions of the pins. For the current model, the pin length was set to 40 mm. In addition, 2T_3 represents a rigid body translation along the length of the telescoping slider mechanism within the body of the fixator. Also, 3T_4 and 1T_2 are pure rotations at the ball and socket joints. The rotation sequence of the ball and socket joints for α_i , β_i , and γ_i , i=1 and 2, is Rot(x)-Rot(y')-Rot(z''), as a standard Briant angle convention[3]. The maximum allowable length of the translation for the telescoping slider and the rotation angle for the ball and socket joints are 40 mm and 20°, respectively. The initial malalignment is expressed within GT_P as rotations around the axes of the fragment and a translation along the long axis of the bone. To realign the bones, the resulting system of equations is solved for the rotation or translation at each joint. For this model the length of the bone from the fracture site to the closest pin on both the proximal and distal segments was set to 80 mm.

After substituting in the known values in the transformation matrices, ${}^{G}T_{D}$, ${}^{D}T_{1}$ and ${}^{4}T_{P}$, Eq. 1 can be reduced as follows:

$${}^{1}\mathbf{T}_{4} = {}^{1}\mathbf{T}_{2} \cdot {}^{2}\mathbf{T}_{3} \cdot {}^{3}\mathbf{T}_{4},$$
 (2),

where

$${}^{1}\mathbf{T}_{4}(1,4) = \cos(\beta_{1}) * t_{2},$$

$${}^{1}\mathbf{T}_{4}(2,4) = -\sin(\alpha_{1})*\cos(\beta_{1})*t_{2},$$

$${}^{1}T_{4}(3,4) = Cos(\alpha_{1}) * Cos(\beta_{1}) * t_{2}$$

where α_1 , β_1 and t_2 are two angle parameters in 1T_2 and the translation parameter in 2T_3 , respectively (Fig. 1). After solving for t_2 , α_1 and β_1 using the three equations in Eq. (2) and calculating γ_1 using the projection method, α_2 , β_2 , and γ_2 are solved using three other equations in Eq. (2). Two clinical cases, a rotational malalignment combined with a fracture gap between two bone segments and a varus angulation, were investigated to validate the model and the analysis involved.

RESULTS

In the first case, a rotational malalignment combined with a fracture gap between two bone segments is considered. The initial gap and the rotational malalignment were set to 6 mm and 30°, respectively. One possible solution is given in Fig. 2 for both the proximal and the distal ball joints as $(\alpha_1, \beta_1, \gamma_1)=(5^\circ, -18^\circ, 16^\circ), (\alpha_2, \gamma_1)=(5^\circ, -18^\circ, 16^\circ)$ β_2 , γ_2)=(-10°, -16°, 16°) and no extension in the slider (Fig. 2). In the second case, 10° varus angulational deformity correction is simulated and visualized (Fig. 3). The candidate solution to correct the deformity was estimated as $(\alpha_1, \beta_1, \gamma_1)=(5^\circ, 0^\circ, 0^\circ)$, $(\alpha_2, \beta_2, \gamma_2)=(5^\circ, 0^\circ, 0^\circ)$ 0°) and $t_2 = 12.3$ mm. The graphical models clearly depict where the adjustments were performed to reduce the bone segment malalignment perfectly. It is important to realize that the component adjustments are sequence dependent according to the order of the transformation matrices multiplication assigned. The order of adjustment can be changed which may produce a different solution, thus making the adjustability of a fixator nonunique.

DISCUSSION

External fixators are often applied to polytrauma cases under an environment in which adequate imaging equipment may not be available to provide acceptable fracture segment reduction. Therefore, subsequent adjustment of the fracture ends is often required for a more complex device. Unfortunately, this issue of design requirement rarely has been investigated and this study provides a systematic approach to this problem. Using the kinematic linkage model, the rotational malalignment combined with a fracture gap displacement could be modeled and analyzed. Although the adjustment solution may not be unique, the model quantified the order and the magnitude of translation and rotation for each joint of the fixator to align the bone These calculated adjustments could reduce the fracture malalignment and angulation, but they may not be feasible due to bone fragments collision or to produce excessive soft tissue stretching during reduction. Visualization of the sequential adjustments used to reduce the fracture gap based on alternative solutions should help to determine the clinical feasibility of the reduction process. Further development of the model should include constraints to avoid collision or excessive soft tissue extension. These graphical model and analysis simulation are powerful tools to evaluate and validate the device performance.

The simulation model allows the kinematic analysis results to be visualized. The bone and pin length, and the initial deformity are input to the model. This model and analysis technique can be used to illustrate the necessary fixator component adjustment to correct the malalignment, making the system a valuable preoperative planning tool and a teaching aid. This simulation technology is now being extended to study the strength and stiffness of the fixator under physiological loading. Stress/strain analyses can also be incorporated in the future to reduce the need for mechanical testing of these devices. By incorporating the translational or rotational limit of each joint, the model can also be used to determine the limits of adjustability inherent to a given fixator design.

REFERENCES

- Chao, E.Y.S., Rim, K., Smidt, G.L. and Johnston, R.C., 1970, "The Application of 4 X 4 Matrix Method to the Correction of the Measurements of Hip Joint Rotations," J. Biomech., Vol. 3, pp. 459-471.
- McCarthy J.M., 1986, Kinematics of Robotic Manipulators, MIT Press
- Nikravesh P.E., 1988, Computer-Aided Analysis of Mech. Sys., Prentice Hall.

ACKNOWLEDGEMENT

This study was supported in part by U.S. Army Medical Research and Material Command(USAMRMC) and a gift from EBI Biomedical System Inc.




Figure 1. A bone and fixator system with rotational malalignment combined with a fracture gap.




Figure 2. A bone and fixator system after correction.




Figure 3 A bone and fixator system in a 10° varus angular deformity correction.